



Ref. Certif. No.

US-41116-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Power Supply

Name and address of the applicant

NIPRON CO LTD
2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095
Japan

Name and address of the manufacturer

NIPRON CO LTD
2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095
Japan

Name and address of the factory

NIPRON CO LTD
3249 OIZU INDUSTRIAL PARK IN OYODO KOGYO DANCHI
OAZA-YAMAOIZU AZA-NAKAJIMA MEIWA-CHO TAKI-GUN,
MIE-KEN, 515-0303
Japan

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

☒ Additional Information on page 2
mOZP-350-12-xyEnz
Input: 100-240 Vac, 4.8 A, 50-60 Hz
Output: 12 Vdc, 25 A (42 Apeak)
☒ Additional Information on page 2

Trademark / Brand (if any)

N/A

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

mOZP-350-12-xyEnz, mOZP-350-15-xyEnz, mOZP-350-24-xyEnz,
mOZP-350-30-xyEnz, mOZP-350-36-xyEnz
☒ Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013,
EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021
National Differences specified in the CB Test Report.

A sample of the product was tested and found to be in conformity with

☐ Additional Information on page 2
IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012,
IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E358786-D1017-1/A0/C0-CB issued on 2022-10-28

This CB Test Certificate is issued by the National Certification Body



- ☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-11-01

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-41116-UL

Factory(ies):

NIPRON CO LTD
2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095
Japan

Additional Model Detail(s):

mOZP-350-12-xyEnz, mOZP-350-15-xyEnz, mOZP-350-24-xyEnz,
mOZP-350-30-xyEnz, mOZP-350-36-xyEnz, mOZP-350-48-xyEnz,
("m": is suffix not affecting safety. The suffix does not affect basic safety or essential performance. x is "J", "T", or "W", y is "0" or
"S", n is any number "0" to "9" or any letter "A" to "Z" except for "E" or blank and z is "-K" or "-C" or blank)

Additional Ratings:

mOZP-350-15-xyEnz
Input: 100-240 Vac, 4.8 A, 50-60 Hz
Output: 15 Vdc, 20 A (40 Apeak)

mOZP-350-24-xyEnz except for mOZP-350-24-JSE3 and mOZP-350-24-WSE3
Input: 100-240 Vac, 5.5 A, 50-60 Hz
Output: 24 Vdc, 14.6 A (25 Apeak)

mOZP-350-24-JSE3, mOZP-350-24-WSE3
Input: 100-240 Vac, 3.1 A, 50-60 Hz
Output: 24 Vdc, 11.67 A (12 Apeak)

mOZP-350-30-xyEnz
Input: 100-240 Vac, 5.5 A, 50-60 Hz
Output: 30 Vdc, 11.7 A (20 Apeak)

mOZP-350-36-xyEnz
Input: 100-240 Vac, 5.5 A, 50-60 Hz
Output: 36 Vdc, 9.8 A (16.7 Apeak)

mOZP-350-48-xyEnz
Input: 100-240 Vac, 5.5 A, 50-60 Hz
Output: 48 Vdc, 7.3 A (12.5 Apeak)

Additional information (if necessary)



☒ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☐ UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2022-11-01

Signature:

Jolanta M. Wroblewska