

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**CB TEST CERTIFICATE**

| | |
|---|--|
| Product | Power Supply |
| Name and address of the applicant | NIPRON CO LTD 2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095 JAPAN |
| Name and address of the manufacturer | NIPRON CO LTD 2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095 JAPAN |
| Name and address of the factory | NIPRON CO LTD 2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095 JAPAN |
| Note: When more than one factory, please report on page 2 | <input checked="" type="checkbox"/> Additional Information on page 2 |
| Ratings and principal characteristics | See Page 2 |
| Trademark (if any) | None |
| Customer's Testing Facility (CTF) Stage used | |
| Model / Type Ref. | mFZP-075-12x, mFZP-075-15x, mFZP-075-24x, mFZP-075-5x (where x maybe maximum 20 characters, any alphanumeric character, hyphen, slash or blank) |
| Additional information (if necessary may also be reported on page 2) | The product has been evaluated with the exception of Risk Management assessment. <input checked="" type="checkbox"/> Additional Information on page 2 |
| A sample of the product was tested and found to be in conformity with | IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005 |
| As shown in the Test Report Ref. No. which forms part of this Certificate | E358786-D1008-1/A0/C0-CB issued on 2021-06-15 |

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-06-24

Signature:



Jolanta M. Wroblewska

Factories:

NIPRON CO LTD
3249 OIZU INDUSTRIAL PARK IN OYODO KOGYO DANCHI OAZA-YAMAOIZU
AZA-NAKAJIMA MEIWA-CHO TAKI-GUN, MIE-KEN, 515-0303
JAPAN

Ratings:

Model mFZP-075-5x:

Input: 100-240 Vac, 50-60 Hz, 1.2-0.8 A
Output: 5 Vdc, 10 A (Peak 15 A)

Model mFZP-075-12x:

Input: 100-240 Vac, 50-60 Hz, 1.5-0.9 A
Output: 12 Vdc, 6.25 A (Peak 12.5 A)

Model mFZP-075-15x:

Input: 100-240 Vac, 50-60 Hz, 1.5-0.9 A
Output: 15 Vdc, 5 A (Peak 10 A)

Model mFZP-075-24x:

Input: 100-240 Vac, 50-60 Hz, 1.5-0.9 A
Output: 24 Vdc, 3.13 A (Peak 6.25 A)

Additionally evaluated to:

EN 60601-1:2006/A1:2013/A12:2014; National Differences specified in the CB Test Report.

Additional information (if necessary)

- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Signature:



Jolanta M. Wroblewska

Date: 2021-06-24