



Ref. Certif. No.

US-18446-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Power Supply

Name and address of the applicant
Nom et adresse du demandeur

NIPRON CO LTD
2-57 OHAMA-CHO
AMAGASAKI-SHI
HYOGO-KEN 660-0095, JAPAN

Name and address of the manufacturer
Nom et adresse du fabricant

NIPRON CO LTD
2-57 OHAMA-CHO
AMAGASAKI-SHI
HYOGO-KEN 660-0095, JAPAN

Name and address of the factory
Nom et adresse de l'usine

NIPRON CO LTD
2-57 OHAMA-CHO AMAGASAKI-SHI
HYOGO-KEN 660-0095
JAPAN

Note: When more than one factory, please report on page 2
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{eme} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Additional Information on page 2
See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

Nipron

Model / Type Ref.
Ref. De type

GPSA-360-12XXXX, GPSA-360-24XXXX, mGPSA-360-12XXXX, and mGPSA-360-24XXXX, See Page 2

Additional information (if necessary may also be reported on page 2)
Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2^{eme} page

National Differences specified in the CB Test Report.
 Additional Information on page 2

A sample of the product was tested and found to be in conformity with
Un échantillon de ce produit a été essayé et a été considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat

E161936-A63-CB-1 issued on 2012-02-06

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-02-07

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-18446-UL

Model Details:

GPSA-360-12XXXX,GPSA-360-24XXXX

mGPSA-360-24XXXX,mGPSA-360-12XXXX (' X ' = Maybe any alphanumeric character, ' - ' or blank, not affected to safety.)

Ratings:

Models GPSA-360-12XXXX and mGPSA-360-12XXXX

Input: 100-240 Vac, 50/60 Hz, 6.3-2.4 A

Output: CH1: 12 Vdc, 30 A; CH2: 12 VSB, 0.3 A (Total maximum power 363.6 W.)

Output Peak: CH1: 12 Vdc, 40 A; CH2: 12 VSB, 0.3 A (Total maximum power 483.6 W.)

Models GPSA-360-24XXXX and mGPSA-360-24XXXX

Input: 100-240 Vac, 50/60 Hz, 6.3-3.0 A

Output: CH1: 24 Vdc, 15 A; CH2: 12 VSB, 0.3 A (Total maximum power 363.6 W.)

Output Peak: CH1: 24 Vdc, 20.8 A; CH2: 12 VSB 0.3 A at input 100 Vac (Total maximum power 502.8 W.)

CH1: 24 Vdc, 25 A; CH2: 12 VSB, 0.3 A at input 200 Vac (Total maximum power 603.6 W.)

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2012-02-07

Signature:

Jolanta M. Wroblewska