



IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Power Supply

Name and address of the applicant
Nom et adresse du demandeur

NIPRON CO LTD
3-1-9 MATABEEAZAKIZAEMONSHINDEN
AMAGASAKI-SHI HYOGO-KEN 660-0000, JAPAN

Name and address of the manufacturer
Nom et adresse du fabricant

NIPRON CO LTD
3-1-9 MATABEEAZAKIZAEMONSHINDEN
AMAGASAKI-SHI HYOGO-KEN 660-0000, JAPAN

Name and address of the factory
Nom et adresse de l'usine

NIPRON CO LTD
3-1-9 MATABEEAZAKIZAEMONSHINDEN
AMAGASAKI-SHI HYOGO-KEN 660-0000, JAPAN

Rating and principal characteristics
Valeurs nominales et caractéristiques principales

For ratings see the second page of this certificate.

Trademark (if any)
Marque de fabrique (si elle existe)

Not applicable

Model / Type Ref.
Ref. de type

GPSA-360-12XXXX, GPSA-360-24XXXX, mGPSA-360-12XXXX,
mGPSA-360-24XXXX ('X' = Any number 0 to 9, any letter A to Z, '-' or blank,
not affected to safety).

Additional information (if necessary)
Information complémentaire (si nécessaire)

This CB Test Report comprises 7 enclosures.

A sample of the product was tested and found
to be in conformity with
*Un échantillon de ce produit a été essayé et a été
considéré conforme à la*

PUBLICATION

**IEC 60601-1 (1988) Second Edition,
with Amendment No. 1 (1991) and No. 2 (1995) with the exception of:**
Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and Clause 52.1,
Programmable Electronic Systems. See Test Report for National Differences.

as shown in the Test Report Ref. No.
which forms part of this Certificate
*comme indiqué dans le Rapport d'essais numéro
de référence qui constitue partie de ce Certificat*

E161936-A33-CB-1

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



**Underwriters
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2009 October 5

Signature:

Jolanta M. Wroblewska

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

For models GPSA-360-12XXXX and mGPSA-360-12XXXX
Input: 100-240 Vac, 50/60 Hz, 6.3-2.4 A
Output: CH1: +12 Vdc, 30 A, CH2: +12 VSB, 0.3 A
(Total maximum power: 363.6 W)
Output peak: CH1: +12 Vdc, 40 A, CH2: +12 VSB, 0.3 A
(Total maximum power: 483.6 W)

For models GPSA-360-24XXXX and mGPSA-360-24XXXX
Input: 100-240 Vac, 50/60 Hz, 6.3-3.0 A
Output: CH1: +24 Vdc, 15 A, CH2: +12 VSB, 0.3 A
(Total maximum power: 363.6 W)
Output peak:
CH1: +24 Vdc, 20.8 A, CH2: +12 VSB, 0.3 A at input 100 Vac
(Total maximum power: 502.8 W)
CH1: +24 Vdc, 25 A, CH2: +12 VSB, 0.3 A at input 200 Vac
(Total maximum power: 603.6 W)

Additional information (if necessary)
Information complémentaire (si nécessaire)

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification




**Underwriters
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008 FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2009 October 5

Signature:



Jolanta M. Wroblewska