

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME
CB TEST CERTIFICATE

Product	Power Supply
Name and address of the applicant	NIPRON CO LTD 2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095 JAPAN
Name and address of the manufacturer	NIPRON CO LTD 2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095 JAPAN
Name and address of the factory	NIPRON CO LTD 2-57 OHAMA-CHO AMAGASAKI-SHI, HYOGO-KEN, 660-0095 JAPAN
Note: When more than one factory, please report on page 2	<input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	See Page 2
Trademark (if any)	N/A
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	mUZP-150-12x, mUZP-150-18x, mUZP-150-24x, mUZP-150-48x <input checked="" type="checkbox"/> Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	The report was revised to include technical modifications. <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E358786-D1004-1/A1/C0-CB issued on 2021-06-10

This CB Test Certificate is issued by the National Certification Body



- ☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-07-06

Original Issue Date: 2017-07-11

Signature:



Jolanta M. Wroblewska

Factories:

NIPRON CO LTD
3249 OAZA-YAMAOIZU, AZA-NAKAJIMA, MEIWA-CHO, TAKI-GUN
MIE-KEN, 515-0303 JAPAN

Additional Model(s) Details:

mUZP-150-12x, mUZP-150-18x, mUZP-150-24x and mUZP-150-48x (where x maybe maximum 20 characters, any alphanumeric character, hyphen, slash or blank)

Ratings:

mUZP-150-12x Input: AC 100 -240 V, 50 - 60 Hz, 3.0 A
Output: 12 V, 12.5 A (33.4 A Peak)
mUZP-150-18x Input: AC 100 -240 V, 50 - 60 Hz, 2.9 A
Output: 18 V, 8.4 A (22.3 A Peak)
mUZP-150-24x Input: AC 100 -240 V, 50 - 60 Hz, 3.2 A
Output: 24 V, 6.3 A (16.7 A Peak)
mUZP-150-48x Input: AC 100 -240 V, 50 - 60 Hz, 3.1 A
Output: 48 V, 3.2 A (8.4 A Peak)

Additionally evaluated to:

EN 60601-1:2006/A1:2013/A12:2014; National Differences specified in the CB Test Report.

The Risk Management requirements of the standard were not addressed.

Summary of Modifications:

The original report was modified to include the following changes/additions:
Addition of an alternate components. See Test Report for additional details.

Additional information (if necessary)



- ☒ UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☐ UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Signature:



Jolanta M. Wroblewska

Date: 2021-07-06

Original Issue Date: 2017-07-11